BIOZOL

Diagnostica Vertrieb GmbH, Oehleckerring 11-13

22419 Hamburg, Germany

Telephone: +49 (0)89 3799666-6 | **Fax:** +49 (0)89 3799666-99

E-Mail: info@biozol.de

Bitte beachten Sie: Dieses Dokument wurde automatisch erstellt und ist kein Ersatz für das Originaldokument des Herstellers.

Product Datasheet

CircuLex Human Chitotriosidase ELISA Kit

MBL-CY-8074

Artikelname	CircuLex Human Chitotriosidase ELISA Kit
Artikelnummer	MBL-CY-8074
Hersteller Artikelnummer	CY-8074
Alternativnummer	MBL-CY-8074
Hersteller	MBL
Kategorie	Kits/Assays
Applikation	ELISA
Spezies Reaktivität	Hu
Produktbeschreibung	CircuLex Human Chitotriosidase ELISA Kit...
Molekulargewicht	Dilution factors need to be taken into consideration in calculating the human chitotriosidase concentration. Results exceeding human chitotriosidase level of 3.6 ng/ml should be repeated with diluted samples.
Detektionsbereich	Dilution factors need to be taken into consideration in calculating the human chitotriosidase concentration. Results exceeding human chitotriosidase level of 3.6 ng/ml should be repeated with diluted samples.
Sensitivitaet	better than 48.3 pg/ml of sample.
Target-Kategorie	CHITO