

Bitte beachten Sie: Dieses Dokument wurde automatisch erstellt und ist kein Ersatz für das Originaldokument des Herstellers.

## Product Datasheet

### **MUC1 / CA15-3 / EMA / CD227 (Epithelial Marker) (MUC1/2278R), CF405S conjugate, 0.1mg/mL, Clone: [MUC1/2278R], Rabbit, Monoclonal BOT-BNC042278-500**

|                          |   |
|--------------------------|---|
| Artikelname              | MUC1 / CA15-3 / EMA / CD227 (Epithelial Marker) (MUC1/2278R), CF405S conjugate, 0.1mg/mL, Clone: [MUC1/2278R], Rabbit, Monoclonal   |
| Artikelnummer            | BOT-BNC042278-500   |
| Hersteller Artikelnummer | BNC042278-500   |
| Alternativnummer         | BOT-BNC042278-500-500UL   |
| Hersteller               | Biotium   |
| Wirt                     | Rabbit  |
| Kategorie                | Antikörper  |
| Applikation              | IHC   |
| Spezies Reaktivität      | Human   |
| Immunogen                | Recombinant full-length human MUC1 protein  |
| Konjugation              | CF405S  |
| Produktbeschreibung      | This MAb reacts with MUC1, a large transmembrane glycoprotein expressed on the ductal surface of normal glandular epithelia. It is used as tracer agent in CA15.3 assays. The extracellular domain of MUC1 largely consists of a highly conserved, O-glyco... |
| Klonalität               | Monoclonal  |
| Konzentration            | 0.1 mg/mL   |

|                        |   |
|------------------------|---|
| Klon-Bezeichnung       | [MUC1/2278R]  |
| Molekulargewicht       | 265-400 kDa   |
| UniProt                | <a href="#">P15941</a>  |
| Puffer                 | PBS, 0.1% BSA, 0.05% azide  |
| Quelle                 | Animal  |
| Anwendungsbeschreibung | Higher concentration may be required for direct detection using primary antibody conjugates than for indirect detection with secondary antibody Immunohistology (formalin): 0.5-1 ug/mL for 30 minutes at RT Staining of formalin-fixed tissues requires boiling tissue sections in 10 mM citrate buffer, pH 6.0, for 10-20 minutes followed by cooling at RT for 20 minutes Optimal dilution for a specific application should be determined by user |