

Bitte beachten Sie: Dieses Dokument wurde automatisch erstellt und ist kein Ersatz für das Originaldokument des Herstellers.

## Product Datasheet

### **TACSTD2 / TROP2 (Epithelial Marker) (TACSTD2/2151), 1mg/mL, Clone: [TACSTD2/2151], Mouse, Monoclonal BOT-BNUM2151-50**

|                          |   |
|--------------------------|---|
| Artikelname              | TACSTD2 / TROP2 (Epithelial Marker) (TACSTD2/2151), 1mg/mL, Clone: [TACSTD2/2151], Mouse, Monoclonal  |
| Artikelnummer            | BOT-BNUM2151-50   |
| Hersteller Artikelnummer | BNUM2151-50   |
| Alternativnummer         | BOT-BNUM2151-50-50UL  |
| Hersteller               | Biotium   |
| Wirt                     | Mouse   |
| Kategorie                | Antikörper  |
| Applikation              | IHC, WB   |
| Spezies Reaktivität      | Human   |
| Immunogen                | Recombinant fragment of human TACSTD2 protein (around aa 31-274) (exact sequence is proprietary)  |
| Produktbeschreibung      | TACSTD2 is a cell surface glycoprotein receptor. It is a single pass type I membrane protein containing one thyroglobulin type-1 domain, an epidermal growth factor-like repeat, a phosphatidylinositol binding site and tyrosine phosphorylation sites ne... |
| Klonalität               | Monoclonal  |
| Konzentration            | 1 mg/mL   |
| Klon-Bezeichnung         | [TACSTD2/2151]  |

|                        |  |
|------------------------|--|
| Molekulargewicht       | 40 kDa   |
| UniProt                | <a href="#">P09758</a>   |
| Puffer                 | PBS, no BSA, no azide  |
| Quelle                 | Animal   |
| Anwendungsbeschreibung | Higher concentration may be required for direct detection using primary antibody conjugates than for indirect detection with secondary antibody Immunohistology (formalin): 1-2 ug/mL for 30 minutes at RT Staining of formalin-fixed tissues requires boiling tissue sections in 10 mM citrate buffer, pH 6.0, for 10-20 minutes followed by cooling at RT for 20 minutes Western blotting 0.5-1 ug/mL Optimal dilution for a specific application should be determined by user |