BIOZOL

Diagnostica Vertrieb GmbH, Oehleckerring 11-13

22419 Hamburg, Germany

Telephone: +49 (0)89 3799666-6 | **Fax:** +49 (0)89 3799666-99

E-Mail: info@biozol.de

Bitte beachten Sie: Dieses Dokument wurde automatisch erstellt und ist kein Ersatz für das Originaldokument des Herstellers.

Product Datasheet

SARS-CoV-2 Triplex PCR kit (Form F)

ATB-89-03

Artikelname	SARS-CoV-2 Triplex PCR kit (Form F)
Artikelnummer	ATB-89-03
Hersteller Artikelnummer	89-03
Alternativnummer	ATB-89-03
Hersteller	Astra Biotech GmbH
Kategorie	Kits/Assays
Spezies Reaktivität	Human
Produktbeschreibung	Text zusatzlich mitgeschickt...
Sensitivitaet	Analytical Sensitivity 1000 copies of SC2 genomic RNA per 1 mL of sample (or ca. 10 copies per reaction) Diagnostic Sensitivity 100% Diagnostic Specificity 100%
Proben	Human samples from nasopharyngeal swabs, oropharyngeal swabs and saliva samples
Target-Kategorie	SARS-CoV-2

Anwendungsbeschreibung	SARS-CoV-2 Triplex PCR kit is a one-step reverse transcription real-time PCR multiplex test for detection of SARS-CoV-2 nucleic acids from nasopharyngeal swabs, oropharyngeal swabs and saliva samples from individuals suspected of COVID-19. The kit simultaneously detects three targets, in ORF1ab, ORF8, and N protein coding regions, each on separate detection channel. The kit was validated to be used with following real-time PCR cyclers: Qiagen Rotor-Gene Q (also can be used with Corbett Research Rotor-Gene 3000/6000), Bio-Rad CFX96 (also can be used with CFX96 Touch), DNA-Technology DTprime (corresponds to DTlite), Sacace Biotechnologies SaCycler, Roche Light Cycler 96 Features: Advanced test design three SARS-CoV-2 specific targets chosen from least variable regions in ORF1ab, ORF8, and N genes multiplicity of targets increases the robustness of the test in case of possible arising mutations in the regions where primers or probes bind Control of the entire diagnostic process endogenous internal control (transcript of one of human housekeeping genes) allows to monitor the quality of sampling, RNA extraction procedure, reverse transcription and PCR amplification, as well as detection reactions positive control (in vitro transcribed RNA of target regions) allows to monitor the quality of reverse transcription and PCR amplification and detection reactions negative control allows to monitor the cases of possible contamination Single tube design simultaneous detection of three targets + internal control in one tube internal control is included in master mix - no additional pipetting required