

Bitte beachten Sie: Dieses Dokument wurde automatisch erstellt und ist kein Ersatz für das Originaldokument des Herstellers.

Product Datasheet

Human CD28(Cluster Of Differentiation 28) Microsample ELISA Kit EBT-ELK0672MS

Artikelname	Human CD28(Cluster Of Differentiation 28) Microsample ELISA Kit
Artikelnummer	EBT-ELK0672MS
Hersteller Artikelnummer	ELK0672MS
Alternativnummer	EBT-ELK0672MS-96, EBT-ELK0672MS-48, EBT-ELK0672MS-96X5
Hersteller	ELK Biotechnology
Kategorie	Kits/Assays
Spezies Reaktivität	Human
Konzentration	50 ng/mL
Detektionsbereich	0.78-50 ng/mL
Sensitivitaet	0.32 ng/mL
UniProt	P10747
Proben	serum, plasma, cell lysates and other biological fluids

Anwendungsbeschreibung

Assay Type: Sandwich. Assay length: 3.5h. Research Area: CD & Adhesion molecule, Tumor immunity, Infection immunity, Immune molecule, . Test principle: The test principle applied in this kit is Sandwich enzyme immunoassay. The microtiter plate provided in this kit has been pre-coated with an antibody specific to Human CD28. Standards or samples are added to the appropriate microtiter plate wells then with a biotin-conjugated antibody specific to Human CD28. Next, Avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. After TMB substrate solution is added, only those wells that contain Human CD28, biotin-conjugated antibody and enzyme-conjugated Avidin will exhibit a change in color. The enzyme-substrate reaction is terminated by the addition of sulphuric acid solution and the color change is measured spectrophotometrically at a wavelength of 450nm 10nm. The concentration of Human CD28 in the samples is then determined by comparing the OD of the samples to the standard curve