

Bitte beachten Sie: Dieses Dokument wurde automatisch erstellt und ist kein Ersatz für das Originaldokument des Herstellers.

Product Datasheet

Dog CTXI(Cross Linked C-Telopeptide Of Type I Collagen) ELISA Kit EBT-ELK10020

Artikelname	Dog CTXI(Cross Linked C-Telopeptide Of Type I Collagen) ELISA Kit
Artikelnummer	EBT-ELK10020
Hersteller Artikelnummer	ELK10020
Alternativnummer	EBT-ELK10020-96, EBT-ELK10020-96X5, EBT-ELK10020-48
Hersteller	ELK Biotechnology
Kategorie	Kits/Assays
Spezies Reaktivität	Canine
Konzentration	10 ng/mL
Detektionsbereich	0.16-10 ng/mL
Sensitivitaet	0.075 ng/mL
Proben	serum, plasma and other biological fluids

Anwendungsbeschreibung

Assay Type: Sandwich. Assay length: 3.5h. Research Area: Metabolic pathway, Bone metabolism,. Test principle: The test principle applied in this kit is Sandwich enzyme immunoassay. The microtiter plate provided in this kit has been pre-coated with an antibody specific to Dog CTXI. Standards or samples are added to the appropriate microtiter plate wells then with a biotin-conjugated antibody specific to Dog CTXI. Next, Avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. After TMB substrate solution is added, only those wells that contain Dog CTXI, biotin-conjugated antibody and enzyme-conjugated Avidin will exhibit a change in color. The enzyme-substrate reaction is terminated by the addition of sulphuric acid solution and the color change is measured spectrophotometrically at a wavelength of 450nm 10nm. The concentration of Dog CTXI in the samples is then determined by comparing the OD of the samples to the standard curve