

Bitte beachten Sie: Dieses Dokument wurde automatisch erstellt und ist kein Ersatz für das Originaldokument des Herstellers.

Product Datasheet

Human GFAP(Glial Fibrillary Acidic Protein) ELISA Kit EBT-ELK1626

Artikelname	Human GFAP(Glial Fibrillary Acidic Protein) ELISA Kit
Artikelnummer	EBT-ELK1626
Hersteller Artikelnummer	ELK1626
Alternativnummer	EBT-ELK1626-96, EBT-ELK1626-48, EBT-ELK1626-96X5
Hersteller	ELK Biotechnology
Kategorie	Kits/Assays
Spezies Reaktivität	Human
Konzentration	10 ng/mL
Detektionsbereich	0.16-10 ng/mL
Sensitivitaet	0.058 ng/mL
UniProt	P14136
Proben	serum, plasma, tissue homogenates, cell lysates, cell culture supernates and other biological fluids

Anwendungsbeschreibung

Assay Type: Sandwich. Assay length: 3.5h. Research Area: Infection immunity, Neuro science,. Test principle: The test principle applied in this kit is Sandwich enzyme immunoassay. The microtiter plate provided in this kit has been pre-coated with an antibody specific to Human GFAP. Standards or samples are added to the appropriate microtiter plate wells then with a biotin-conjugated antibody specific to Human GFAP. Next, Avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. After TMB substrate solution is added, only those wells that contain Human GFAP, biotin-conjugated antibody and enzyme-conjugated Avidin will exhibit a change in color. The enzyme-substrate reaction is terminated by the addition of sulphuric acid solution and the color change is measured spectrophotometrically at a wavelength of 450nm 10nm. The concentration of Human GFAP in the samples is then determined by comparing the OD of the samples to the standard curve