

Bitte beachten Sie: Dieses Dokument wurde automatisch erstellt und ist kein Ersatz für das Originaldokument des Herstellers.

Product Datasheet

Human CD34(Cluster Of Differentiation 34) Microsample ELISA Kit EBT-ELK2707MS

Artikelname	Human CD34(Cluster Of Differentiation 34) Microsample ELISA Kit
Artikelnummer	EBT-ELK2707MS
Hersteller Artikelnummer	ELK2707MS
Alternativnummer	EBT-ELK2707MS-96, EBT-ELK2707MS-48, EBT-ELK2707MS-96X5
Hersteller	ELK Biotechnology
Kategorie	Kits/Assays
Spezies Reaktivität	Human
Konzentration	20 ng/mL
Detektionsbereich	0.32-20 ng/mL
Sensitivitaet	0.118 ng/mL
UniProt	P28906
Proben	tissue homogenates, cell lysates, cell culture supernates and other biological fluids

Anwendungsbeschreibung	<p>Assay Type: Sandwich. Assay length: 3.5h. Research Area: CD & Adhesion molecule, Developmental science, Genetic science,. Test principle: The test principle applied in this kit is Sandwich enzyme immunoassay. The microtiter plate provided in this kit has been pre-coated with an antibody specific to Human CD34. Standards or samples are added to the appropriate microtiter plate wells then with a biotin-conjugated antibody specific to Human CD34. Next, Avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. After TMB substrate solution is added, only those wells that contain Human CD34, biotin-conjugated antibody and enzyme-conjugated Avidin will exhibit a change in color. The enzyme-substrate reaction is terminated by the addition of sulphuric acid solution and the color change is measured spectrophotometrically at a wavelength of 450nm 10nm. The concentration of Human CD34 in the samples is then determined by comparing the OD of the samples to the standard curve</p>
------------------------	--