

Bitte beachten Sie: Dieses Dokument wurde automatisch erstellt und ist kein Ersatz für das Originaldokument des Herstellers.

Product Datasheet

Human COL17(Collagen Type XVII) Microsample ELISA Kit EBT-ELK2988MS

Artikelname	Human COL17(Collagen Type XVII) Microsample ELISA Kit
Artikelnummer	EBT-ELK2988MS
Hersteller Artikelnummer	ELK2988MS
Alternativnummer	EBT-ELK2988MS-96, EBT-ELK2988MS-48, EBT-ELK2988MS-96X5
Hersteller	ELK Biotechnology
Kategorie	Kits/Assays
Spezies Reaktivität	Human
Konzentration	200 ng/mL
Detektionsbereich	3.13-200 ng/mL
Sensitivitaet	1.1 ng/mL
Proben	serum, plasma, tissue homogenates and other biological fluids

Anwendungsbeschreibung

Assay Type: Sandwich. Assay length: 3.5h. Research Area: Signal transduction, Metabolic pathway, Developmental science, Nutrition metabolism, Dermatology, . Test principle: The test principle applied in this kit is Sandwich enzyme immunoassay. The microtiter plate provided in this kit has been pre-coated with an antibody specific to Human COL17. Standards or samples are added to the appropriate microtiter plate wells then with a biotin-conjugated antibody specific to Human COL17. Next, Avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. After TMB substrate solution is added, only those wells that contain Human COL17, biotin-conjugated antibody and enzyme-conjugated Avidin will exhibit a change in color. The enzyme-substrate reaction is terminated by the addition of sulphuric acid solution and the color change is measured spectrophotometrically at a wavelength of 450nm 10nm. The concentration of Human COL17 in the samples is then determined by comparing the OD of the samples to the standard curve