

Bitte beachten Sie: Dieses Dokument wurde automatisch erstellt und ist kein Ersatz für das Originaldokument des Herstellers.

Product Datasheet

Human FcαR(Fc Fragment Of IgA Receptor) Microsample ELISA Kit EBT-ELK3381MS

Artikelname	Human FcαR(Fc Fragment Of IgA Receptor) Microsample ELISA Kit
Artikelnummer	EBT-ELK3381MS
Hersteller Artikelnummer	ELK3381MS
Alternativnummer	EBT-ELK3381MS-96, EBT-ELK3381MS-48, EBT-ELK3381MS-96X5
Hersteller	ELK Biotechnology
Kategorie	Kits/Assays
Spezies Reaktivität	Human
Konzentration	50 ng/mL
Detektionsbereich	0.79-50 ng/mL
Sensitivitaet	0.37 ng/mL
UniProt	P24071
Proben	serum, plasma, tissue homogenates and other biological fluids

Anwendungsbeschreibung

Assay Type: Sandwich. Assay length: 3.5h. Research Area: Signal transduction, CD & Adhesion molecule, Infection immunity,. Test principle: The test principle applied in this kit is Sandwich enzyme immunoassay. The microtiter plate provided in this kit has been pre-coated with an antibody specific to Human FcαR. Standards or samples are added to the appropriate microtiter plate wells then with a biotin-conjugated antibody specific to Human FcαR. Next, Avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. After TMB substrate solution is added, only those wells that contain Human FcαR, biotin-conjugated antibody and enzyme-conjugated Avidin will exhibit a change in color. The enzyme-substrate reaction is terminated by the addition of sulphuric acid solution and the color change is measured spectrophotometrically at a wavelength of 450nm ± 10nm. The concentration of Human FcαR in the samples is then determined by comparing the OD of the samples to the standard curve