

Bitte beachten Sie: Dieses Dokument wurde automatisch erstellt und ist kein Ersatz für das Originaldokument des Herstellers.

## Product Datasheet

### Human CDK8(Cyclin Dependent Kinase 8) ELISA Kit EBT-ELK4302

Artikelname	Human CDK8(Cyclin Dependent Kinase 8) ELISA Kit
Artikelnummer	EBT-ELK4302
Hersteller Artikelnummer	ELK4302
Alternativnummer	EBT-ELK4302-96, EBT-ELK4302-48, EBT-ELK4302-96X5
Hersteller	ELK Biotechnology
Kategorie	Kits/Assays
Spezies Reaktivität	Human
Konzentration	20 ng/mL
Detektionsbereich	0.32-20 ng/mL
Sensitivitaet	0.125 ng/mL
UniProt	<a href="#">P49336</a>
Proben	Tissue homogenates and other biological fluids.

Anwendungsbeschreibung

Assay Type: Sandwich. Assay length: 3.5h. Research Area: Tumor immunity,. Test principle: The test principle applied in this kit is Sandwich enzyme immunoassay. The microtiter plate provided in this kit has been pre-coated with an antibody specific to Human CDK8. Standards or samples are added to the appropriate microtiter plate wells then with a biotin-conjugated antibody specific to Human CDK8. Next, Avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. After TMB substrate solution is added, only those wells that contain Human CDK8, biotin-conjugated antibody and enzyme-conjugated Avidin will exhibit a change in color. The enzyme-substrate reaction is terminated by the addition of sulphuric acid solution and the color change is measured spectrophotometrically at a wavelength of 450nm 10nm. The concentration of Human CDK8 in the samples is then determined by comparing the OD of the samples to the standard curve