

Bitte beachten Sie: Dieses Dokument wurde automatisch erstellt und ist kein Ersatz für das Originaldokument des Herstellers.

## Product Datasheet

### Mouse CD163(Cluster Of Differentiation) ELISA Kit EBT-ELK4483

|                          |   |
|--------------------------|---|
| Artikelname              | Mouse CD163(Cluster Of Differentiation) ELISA Kit |
| Artikelnummer            | EBT-ELK4483                                       |
| Hersteller Artikelnummer | ELK4483   |
| Alternativnummer         | EBT-ELK4483-96, EBT-ELK4483-48, EBT-ELK4483-96X5  |
| Hersteller               | ELK Biotechnology                                 |
| Kategorie                | Kits/Assays                                       |
| Spezies Reaktivität      | Mouse   |
| Konzentration            | 2000 pg/mL  |
| Detektionsbereich        | 31.25-2000 pg/mL                                  |
| Sensitivitaet            | 13.8 pg/mL  |
| UniProt                  | <a href="#">Q2VLH6</a>                            |
| Proben                   | Serum, plasma and other biological fluids         |

|                        |   |
|------------------------|---|
| Anwendungsbeschreibung | <p>Assay Type: Sandwich. Assay length: 3.5h. Research Area: CD &amp; Adhesion molecule, Metabolic pathway,. Test principle: The test principle applied in this kit is Sandwich enzyme immunoassay. The microtiter plate provided in this kit has been pre-coated with an antibody specific to Mouse CD163. Standards or samples are added to the appropriate microtiter plate wells then with a biotin-conjugated antibody specific to Mouse CD163. Next, Avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. After TMB substrate solution is added, only those wells that contain Mouse CD163, biotin-conjugated antibody and enzyme-conjugated Avidin will exhibit a change in color. The enzyme-substrate reaction is terminated by the addition of sulphuric acid solution and the color change is measured spectrophotometrically at a wavelength of 450nm 10nm. The concentration of Mouse CD163 in the samples is then determined by comparing the OD of the samples to the standard curve</p> |
|------------------------|---|