

Bitte beachten Sie: Dieses Dokument wurde automatisch erstellt und ist kein Ersatz für das Originaldokument des Herstellers.

Product Datasheet

Human UTI(Urinary Trypsin Inhibitor) ELISA Kit EBT-ELK4874

Artikelname	Human UTI(Urinary Trypsin Inhibitor) ELISA Kit
Artikelnummer	EBT-ELK4874
Hersteller Artikelnummer	ELK4874
Alternativnummer	EBT-ELK4874-96, EBT-ELK4874-48, EBT-ELK4874-96X5
Hersteller	ELK Biotechnology
Kategorie	Kits/Assays
Spezies Reaktivität	Human
Konzentration	30 ng/mL
Detektionsbereich	0.47-30 ng/mL
Sensitivitaet	0.23 ng/mL
Proben	Serum, plasma, urine, tissue homogenates, cell culture supernates and other biological fluids

Anwendungsbeschreibung

Assay Type: Sandwich. Assay length: 3.5h. Research Area: Enzyme & Kinase, Infection immunity, Rheumatology, . Test principle: The test principle applied in this kit is Sandwich enzyme immunoassay. The microtiter plate provided in this kit has been pre-coated with an antibody specific to Human UTI. Standards or samples are added to the appropriate microtiter plate wells then with a biotin-conjugated antibody specific to Human UTI. Next, Avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. After TMB substrate solution is added, only those wells that contain Human UTI, biotin-conjugated antibody and enzyme-conjugated Avidin will exhibit a change in color. The enzyme-substrate reaction is terminated by the addition of sulphuric acid solution and the color change is measured spectrophotometrically at a wavelength of 450nm 10nm. The concentration of Human UTI in the samples is then determined by comparing the OD of the samples to the standard curve