

Bitte beachten Sie: Dieses Dokument wurde automatisch erstellt und ist kein Ersatz für das Originaldokument des Herstellers.

## Product Datasheet

### Human CD8(Cluster of Differentiation 8) Microsample ELISA Kit EBT-ELK5104MS

Artikelname	Human CD8(Cluster of Differentiation 8) Microsample ELISA Kit
Artikelnummer	EBT-ELK5104MS
Hersteller Artikelnummer	ELK5104MS
Alternativnummer	EBT-ELK5104MS-96, EBT-ELK5104MS-48, EBT-ELK5104MS-96X5
Hersteller	ELK Biotechnology
Kategorie	Kits/Assays
Spezies Reaktivität	Human
Konzentration	100 ng/mL
Detektionsbereich	1.57-100 ng/mL
Sensitivitaet	0.62 ng/mL
UniProt	<a href="#">P01732</a>
Proben	Tissue homogenates, cell lysates, cell culture supernates and other biological fluids.

Anwendungsbeschreibung

Assay Type: Sandwich. Assay length: 3.5h. Research Area: Signal transduction, CD & Adhesion molecule, Tumor immunity, Infection immunity, Immune molecule, . Test principle: The test principle applied in this kit is Sandwich enzyme immunoassay. The microtiter plate provided in this kit has been pre-coated with an antibody specific to Human CD8. Standards or samples are added to the appropriate microtiter plate wells then with a biotin-conjugated antibody specific to Human CD8. Next, Avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. After TMB substrate solution is added, only those wells that contain Human CD8, biotin-conjugated antibody and enzyme-conjugated Avidin will exhibit a change in color. The enzyme-substrate reaction is terminated by the addition of sulphuric acid solution and the color change is measured spectrophotometrically at a wavelength of 450nm 10nm. The concentration of Human CD8 in the samples is then determined by comparing the OD of the samples to the standard curve