

Bitte beachten Sie: Dieses Dokument wurde automatisch erstellt und ist kein Ersatz für das Originaldokument des Herstellers.

Product Datasheet

Human LAB7-1(B-Lymphocyte Activation Antigen B7-1) ELISA Kit EBT-ELK5326

| | |
|--------------------------|---|
| Artikelname | Human LAB7-1(B-Lymphocyte Activation Antigen B7-1) ELISA Kit |
| Artikelnummer | EBT-ELK5326 |
| Hersteller Artikelnummer | ELK5326 |
| Alternativnummer | EBT-ELK5326-96, EBT-ELK5326-48, EBT-ELK5326-96X5 |
| Hersteller | ELK Biotechnology |
| Kategorie | Kits/Assays |
| Spezies Reaktivität | Human |
| Konzentration | 5000 pg/mL |
| Detektionsbereich | 78.13-5000 pg/mL |
| Sensitivitaet | 32 pg/mL |
| Proben | Serum, plasma, tissue homogenates and other biological fluids |

Anwendungsbeschreibung

Assay Type: Sandwich. Assay length: 3.5h. Research Area: CD & Adhesion molecule, Tumor immunity, Infection immunity, Immune molecule, Autoimmunity, . Test principle: The test principle applied in this kit is Sandwich enzyme immunoassay. The microtiter plate provided in this kit has been pre-coated with an antibody specific to Human LAB7-1. Standards or samples are added to the appropriate microtiter plate wells then with a biotin-conjugated antibody specific to Human LAB7-1. Next, Avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. After TMB substrate solution is added, only those wells that contain Human LAB7-1, biotin-conjugated antibody and enzyme-conjugated Avidin will exhibit a change in color. The enzyme-substrate reaction is terminated by the addition of sulphuric acid solution and the color change is measured spectrophotometrically at a wavelength of 450nm 10nm. The concentration of Human LAB7-1 in the samples is then determined by comparing the OD of the samples to the standard curve