

Bitte beachten Sie: Dieses Dokument wurde automatisch erstellt und ist kein Ersatz für das Originaldokument des Herstellers.

## Product Datasheet

### EliKine(TM) Human MMP-9 ELISA Kit ASC-KTE6027

|                          |   |
|--------------------------|---|
| Artikelname              | EliKine(TM) Human MMP-9 ELISA Kit   |
| Artikelnummer            | ASC-KTE6027   |
| Hersteller Artikelnummer | KTE6027   |
| Alternativnummer         | ASC-KTE6027-48, ASC-KTE6027-96  |
| Hersteller               | Abbkine Scientific  |
| Kategorie                | Kits/Assays   |
| Applikation              | ELISA   |
| Spezies Reaktivität      | Human   |
| Produktbeschreibung      | EliKine(TM) Human MMP-9 ELISA Kit employs a two-site sandwich ELISA to quantitate Human MMP-9.... |
| Detektionsbereich        | 31.25 pg/mL- 2000 pg/mL   |
| Sensitivitaet            | 16 pg/mL  |
| Tag                      | MMP9  |
| NCBI                     | <a href="#">4318</a>  |
| UniProt                  | <a href="#">P14780</a>  |
| Proben                   | Cell culture supernatantsSerumPlasmaOther biological fluids                                       |

Anwendungsbeschreibung

EliKine(TM) Human MMP-9 ELISA Kit employs a two-site sandwich ELISA to quantitate MMP-9 in samples. An antibody specific for MMP-9 has been pre-coated onto a microplate. Standards and samples are pipetted into the wells and any MMP-9 present is bound by the immobilized antibody. After removing any unbound substances, a biotin-conjugated antibody specific for MMP-9 is added to the wells. After washing, proprietary EliKine(TM) Streptavidin-HRP conjugates is added to the wells. Following a wash to remove any unbound streptavidin-enzyme reagent, a substrate solution is added to the wells and color develops in proportion to the amount of MMP-9 bound in the initial step. The color development is stopped and the intensity of the color is measured.