

Bitte beachten Sie: Dieses Dokument wurde automatisch erstellt und ist kein Ersatz für das Originaldokument des Herstellers.

Product Datasheet

Human Metallothionein-1H (MT1H) ELISA Kit ASC-KTE61510

| | |
|--------------------------|---|
| Artikelname | Human Metallothionein-1H (MT1H) ELISA Kit |
| Artikelnummer | ASC-KTE61510 |
| Hersteller Artikelnummer | KTE61510 |
| Alternativnummer | ASC-KTE61510-48, ASC-KTE61510-96 |
| Hersteller | Abbkine Scientific |
| Kategorie | Kits/Assays |
| Applikation | ELISA |
| Spezies Reaktivität | Human |
| Produktbeschreibung | This Human Metallothionein-1H (MT1H) ELISA Kit employs a two-site sandwich ELISA to quantitate MT1H.... |
| Detektionsbereich | Please inquire |
| Sensitivitaet | Please inquire |
| Tag | MT1H |
| NCBI | 4496 |
| UniProt | P80294 |
| Proben | Cell culture supernatantsSerumPlasmaOther biological fluids |

Anwendungsbeschreibung

This Human Metallothionein-1H (MT1H) ELISA Kit employs a two-site sandwich ELISA to quantitate MT1H in samples. An antibody specific for MT1H has been pre-coated onto a microplate. Standards and samples are pipetted into the wells and any MT1H present is bound by the immobilized antibody. After removing any unbound substances, a biotin-conjugated antibody specific for MT1H is added to the wells. After washing, Streptavidin conjugated Horseradish Peroxidase (HRP) is added to the wells. Following a wash to remove any unbound avidin-enzyme reagent, a substrate solution is added to the wells and color develops in proportion to the amount of MT1H bound in the initial step. The color development is stopped and the intensity of the color is measured.