

Bitte beachten Sie: Dieses Dokument wurde automatisch erstellt und ist kein Ersatz für das Originaldokument des Herstellers.

Product Datasheet

Human N-terminal Telopeptide of Collagen Type-I (NTX) ELISA Kit ASC-KTE62838

| | |
|--------------------------|--|
| Artikelname | Human N-terminal Telopeptide of Collagen Type-I (NTX) ELISA Kit |
| Artikelnummer | ASC-KTE62838 |
| Hersteller Artikelnummer | KTE62838 |
| Alternativnummer | ASC-KTE62838-48, ASC-KTE62838-96 |
| Hersteller | Abbkine Scientific |
| Kategorie | Kits/Assays |
| Applikation | ELISA |
| Spezies Reaktivität | Human |
| Produktbeschreibung | Human N-terminal Telopeptide of Collagen Type-I (NTX) ELISA Kit employs a two-site sandwich ELISA to quantitate NTX in samples.... |
| Detektionsbereich | 1.5 nmol/L-24 nmol/L |
| Sensitivitaet | 0.1 nmol/L |
| Tag | NTX |
| Proben | Cell culture supernatantsSerumPlasmaOther biological fluids |

Anwendungsbeschreibung

Human N-terminal Telopeptide of Collagen Type-I (NTX) ELISA Kit employs a two-site sandwich ELISA to quantitate NTX in samples. An antibody specific for NTX has been pre-coated onto a microplate. Standards and samples are pipetted into the wells and any NTX present is bound by the immobilized antibody. After removing any unbound substances, HRP-Conjugate Human NTX detection antibody is added to the wells. Following a wash to remove any unbound HRP reagent, a Chromogen solution is added to the wells and color develops in proportion to the amount of NTX bound in the initial step. The color development is stopped and the intensity of the color is measured.