

Bitte beachten Sie: Dieses Dokument wurde automatisch erstellt und ist kein Ersatz für das Originaldokument des Herstellers.

Product Datasheet

Rat Follicle Stimulating Hormone (FSH) Microsample Fast ELISA Kit BYT-ORB1806609

| | |
|--------------------------|--|
| Artikelname | Rat Follicle Stimulating Hormone (FSH) Microsample Fast ELISA Kit |
| Artikelnummer | BYT-ORB1806609 |
| Hersteller Artikelnummer | orb1806609 |
| Alternativnummer | BYT-ORB1806609-48,BYT-ORB1806609-96 |
| Hersteller | Biorbyt |
| Kategorie | Kits/Assays |
| Spezies Reaktivität | Rat |
| Produktbeschreibung | Rat Follicle Stimulating Hormone (FSH) Microsample Fast ELISA Kit... |
| Detektionsbereich | 3.13-200ng/mL |
| Sensitivitaet | 1.88 ng/mL |
| UniProt | P18427 |
| Formulierung | Ready to use |
| Proben | serum, plasma, Tissue homogenate and Other biological samples,Sample volume:25µL |
| Target-Kategorie | FSH |

Anwendungsbeschreibung

Application Notes: This ELISA kit uses the Sandwich-ELISA principle. The micro ELISA plate provided in this kit has been pre-coated with an antibody specific to Rat FSH. Samples (or Standards) are added to the micro ELISA plate wells and combined with the specific antibody and biotinylated detection antibody specific for Rat FSH. Then Avidin-Horseradish Peroxidase (HRP) conjugate are added successively to each micro plate well and incubated. Free components are washed away. The substrate solution is added to each well. Only those wells that contain Rat FSH, biotinylated detection antibody and Avidin-HRP conjugate will appear blue in color. The enzyme-substrate reaction is terminated by the addition of stop solution and the color turns yellow. The optical density (OD) is measured spectrophotometrically at a wavelength of 450 nm. The OD value is proportional to the concentration of Rat FSH. You can calculate the concentration of Rat FSH in the samples by comparing the OD of the samples to the standard curve