

Bitte beachten Sie: Dieses Dokument wurde automatisch erstellt und ist kein Ersatz für das Originaldokument des Herstellers.

Product Datasheet

Human Superoxide Dismutase 2, Mitochondrial (SOD2) ELISA Kit BYT-ORB1807204

Artikelname	Human Superoxide Dismutase 2, Mitochondrial (SOD2) ELISA Kit
Artikelnummer	BYT-ORB1807204
Hersteller Artikelnummer	orb1807204
Alternativnummer	BYT-ORB1807204-48,BYT-ORB1807204-96
Hersteller	Biorbyt
Kategorie	Kits/Assays
Spezies Reaktivität	Human
Produktbeschreibung	Human Superoxide Dismutase 2, Mitochondrial (SOD2) ELISA Kit...
Detektionsbereich	0.31-20ng/mL
Sensitivitaet	0.16 ng/mL
UniProt	P04179
Proben	Tissue homogenate and Cell lysate
Target-Kategorie	SOD2

Anwendungsbeschreibung

Application Notes: This ELISA kit uses the Competitive-ELISA principle. The micro ELISA plate provided in this kit has been pre-coated with Human SOD2. During the reaction, Human SOD2 in the sample or standard competes with a fixed amount of Human SOD2 on the solid phase supporter for sites on the Biotinylated Detection Ab specific to Human SOD2. Excess conjugate and unbound sample or standard are washed away, and Avidin-Horseradish Peroxidase (HRP) conjugate are added to each micro plate well and incubated. Then a TMB substrate solution is added to each well. The enzyme-substrate reaction is terminated by the addition of stop solution and the color turns from blue to yellow. The optical density (OD) is measured spectrophotometrically at a wavelength of 450 nm \pm 2 nm. The concentration of Human SOD2 in tested samples can be calculated by comparing the OD of the samples to the standard curve