

Bitte beachten Sie: Dieses Dokument wurde automatisch erstellt und ist kein Ersatz für das Originaldokument des Herstellers.

## Product Datasheet

### Human Hyaluronan Synthase 3 (HAS3) ELISA Kit BYT-ORB1994907

Artikelname	Human Hyaluronan Synthase 3 (HAS3) ELISA Kit
Artikelnummer	BYT-ORB1994907
Hersteller Artikelnummer	orb1994907
Alternativnummer	BYT-ORB1994907-48, BYT-ORB1994907-96
Hersteller	Biorbyt
Kategorie	Kits/Assays
Spezies Reaktivität	Human
Produktbeschreibung	The test principle applied in this kit is Sandwich enzyme immunoassay. The microtiter plate provided in this kit has been pre-coated with an antibody specific to Human HAS3. Standards or samples are added to the appropriate microtiter plate wells the...
Konzentration	5000 pg/mL
Detektionsbereich	78.13-5000 pg/mL
Sensitivitaet	26.3 pg/mL
UniProt	<a href="#">000219</a>
Proben	Serum, plasma, tissue homogenates and other biological fluids

Anwendungsbeschreibung

Application Notes: standard: 5000 pg/mL. Test principle: The test principle applied in this kit is Sandwich enzyme immunoassay. The microtiter plate provided in this kit has been pre-coated with an antibody specific to Human HAS3. Standards or samples are added to the appropriate microtiter plate wells then with a biotin-conjugated antibody specific to Human HAS3. Next, Avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. After TMB substrate solution is added, only those wells that contain Human HAS3, biotin-conjugated antibody and enzyme-conjugated Avidin will exhibit a change in color. The enzyme-substrate reaction is terminated by the addition of sulphuric acid solution and the color change is measured spectrophotometrically at a wavelength of 450nm 10nm. The concentration of Human HAS3 in the samples is then determined by comparing the OD of the samples to the standard curve