

Bitte beachten Sie: Dieses Dokument wurde automatisch erstellt und ist kein Ersatz für das Originaldokument des Herstellers.

Product Datasheet

Human Mitochondrial import receptor subunit TOM6 homolog (TOMM6) ELISA Kit BYT-ORB3145887

| | |
|--------------------------|---|
| Artikelname | Human Mitochondrial import receptor subunit TOM6 homolog (TOMM6) ELISA Kit |
| Artikelnummer | BYT-ORB3145887 |
| Hersteller Artikelnummer | orb3145887 |
| Alternativnummer | BYT-ORB3145887-48, BYT-ORB3145887-5, BYT-ORB3145887-50, BYT-ORB3145887-96 |
| Hersteller | Biorbyt |
| Kategorie | Kits/Assays |
| Applikation | ELISA |
| Spezies Reaktivität | Human |
| Produktbeschreibung | This Human Mitochondrial import receptor subunit TOM6 homolog (TOMM6) ELISA Kit employs a two-site sandwich ELISA to quantitate TOMM6.... |
| Detektionsbereich | Please inquire |
| Sensitivitaet | Please inquire |
| Proben | Cell culture supernatantsSerumPlasmaOther biological fluids |

Anwendungsbeschreibung

Application Notes: This Human Mitochondrial import receptor subunit TOM6 homolog (TOMM6) ELISA Kit employs a two-site sandwich ELISA to quantitate TOMM6 in samples. An antibody specific for TOMM6 has been pre-coated onto a microplate. Standards and samples are pipetted into the wells and any TOMM6 present is bound by the immobilized antibody. After removing any unbound substances, a biotin-conjugated antibody specific for TOMM6 is added to the wells. After washing, Streptavidin conjugated Horseradish Peroxidase (HRP) is added to the wells. Following a wash to remove any unbound avidin-enzyme reagent, a substrate solution is added to the wells and color develops in proportion to the amount of TOMM6 bound in the initial step. The color development is stopped and the intensity of the color is measured