

Bitte beachten Sie: Dieses Dokument wurde automatisch erstellt und ist kein Ersatz für das Originaldokument des Herstellers.

Product Datasheet

Human Urokinase-type Plasminogen Activator Receptor (PLAUR/uPAR) ELISA Kit BYT-ORB3146931

| | |
|--------------------------|--|
| Artikelname | Human Urokinase-type Plasminogen Activator Receptor (PLAUR/uPAR) ELISA Kit |
| Artikelnummer | BYT-ORB3146931 |
| Hersteller Artikelnummer | orb3146931 |
| Alternativnummer | BYT-ORB3146931-48, BYT-ORB3146931-5, BYT-ORB3146931-50, BYT-ORB3146931-96 |
| Hersteller | Biorbyt |
| Kategorie | Kits/Assays |
| Applikation | ELISA |
| Spezies Reaktivität | Human |
| Produktbeschreibung | This Human Urokinase-type Plasminogen Activator Receptor (PLAUR/uPAR) ELISA Kit employs a two-site sandwich ELISA to quantitate PLAUR/uPAR.... |
| Detektionsbereich | Please inquire |
| Sensitivitaet | Please inquire |
| Proben | Cell culture supernatantsSerumPlasmaOther biological fluids |

Anwendungsbeschreibung

Application Notes: This Human Urokinase-type Plasminogen Activator Receptor (PLAUR/uPAR) ELISA Kit employs a two-site sandwich ELISA to quantitate PLAUR/uPAR in samples. An antibody specific for PLAUR/uPAR has been pre-coated onto a microplate. Standards and samples are pipetted into the wells and any PLAUR/uPAR present is bound by the immobilized antibody. After removing any unbound substances, a biotin-conjugated antibody specific for PLAUR/uPAR is added to the wells. After washing, Streptavidin conjugated Horseradish Peroxidase (HRP) is added to the wells. Following a wash to remove any unbound avidin-enzyme reagent, a substrate solution is added to the wells and color develops in proportion to the amount of PLAUR/uPAR bound in the initial step. The color development is stopped and the intensity of the color is measured