

Bitte beachten Sie: Dieses Dokument wurde automatisch erstellt und ist kein Ersatz für das Originaldokument des Herstellers.

Product Datasheet

Human ucOC(Undercarboxylated Osteocalcin) Microsample ELISA Kit BYT-ORB2997326

Artikelname	Human ucOC(Undercarboxylated Osteocalcin) Microsample ELISA Kit
Artikelnummer	BYT-ORB2997326
Hersteller Artikelnummer	orb2997326
Alternativnummer	BYT-ORB2997326-48, BYT-ORB2997326-96
Hersteller	Biorbyt
Kategorie	Kits/Assays
Spezies Reaktivität	Human
Produktbeschreibung	Human ucOC(Undercarboxylated Osteocalcin) Microsample ELISA Kit...
Detektionsbereich	0.32-20 ng/mL
Sensitivitaet	0.111 ng/mL
Proben	serum, plasma, tissue homogenates

Anwendungsbeschreibung

Application Notes: The test principle applied in this kit is Sandwich enzyme immunoassay. The microtiter plate provided in this kit has been pre-coated with an antibody specific to Human ucOC. Standards or samples are added to the appropriate microtiter plate wells then with a biotin-conjugated antibody specific to Human ucOC. Next, Avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. After TMB substrate solution is added, only those wells that contain Human ucOC, biotin-conjugated antibody and enzyme-conjugated Avidin will exhibit a change in color. The enzyme-substrate reaction is terminated by the addition of sulphuric acid solution and the color change is measured spectrophotometrically at a wavelength of 450nm 10nm. The concentration of Human ucOC in the samples is then determined by comparing the OD of the samples to the standard curve