

Bitte beachten Sie: Dieses Dokument wurde automatisch erstellt und ist kein Ersatz für das Originaldokument des Herstellers.

## Product Datasheet

### Human Coagulation Factor XI (F11) ELISA Kit BYT-ORB776081

Artikelname	Human Coagulation Factor XI (F11) ELISA Kit
Artikelnummer	BYT-ORB776081
Hersteller Artikelnummer	orb776081
Alternativnummer	BYT-ORB776081-48, BYT-ORB776081-96
Hersteller	Biorbyt
Kategorie	Kits/Assays
Applikation	ELISA
Spezies Reaktivität	Human
Produktbeschreibung	The test principle applied in this kit is Sandwich enzyme immunoassay. The microtiter plate provided in this kit has been pre-coated with an antibody specific to Coagulation Factor XI(F11)....
Konzentration	40 ng/mL
Detektionsbereich	0.63-40 ng/mL
Sensitivitaet	0.216 ng/mL
UniProt	<a href="#">P03951</a>
Proben	Plasma.

Anwendungsbeschreibung

Application Notes: standard: 40 ng/mL. Test principle: The test principle applied in this kit is Sandwich enzyme immunoassay. The microtiter plate provided in this kit has been pre-coated with an antibody specific to Human F11. Standards or samples are added to the appropriate microtiter plate wells then with a biotin-conjugated antibody specific to Human F11. Next, Avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. After TMB substrate solution is added, only those wells that contain Human F11, biotin-conjugated antibody and enzyme-conjugated Avidin will exhibit a change in color. The enzyme-substrate reaction is terminated by the addition of sulphuric acid solution and the color change is measured spectrophotometrically at a wavelength of 450nm 10nm. The concentration of Human F11 in the samples is then determined by comparing the OD of the samples to the standard curve