

Bitte beachten Sie: Dieses Dokument wurde automatisch erstellt und ist kein Ersatz für das Originaldokument des Herstellers.

Product Datasheet

Mouse Cross Linked C-Telopeptide of Type I Collagen (CTXI) ELISA Kit BYT-ORB776087

Artikelname	Mouse Cross Linked C-Telopeptide of Type I Collagen (CTXI) ELISA Kit
Artikelnummer	BYT-ORB776087
Hersteller Artikelnummer	orb776087
Alternativnummer	BYT-ORB776087-48, BYT-ORB776087-96
Hersteller	Biorbyt
Kategorie	Kits/Assays
Applikation	ELISA
Spezies Reaktivität	Mouse
Produktbeschreibung	This assay employs the competitive inhibition enzyme immunoassay technique. The microtiter plate provided in this kit has been pre-coated with Cross Linked C-Telopeptide Of Type I Collagen (CTXI) protein. Standards or samples are added to the appropri...
Konzentration	10000 pg/mL
Detektionsbereich	156.25-10000 pg/mL
Sensitivitaet	47.5 pg/mL
Proben	serum, plasma, cell lysates, cell culture supernates and other biological fluids

Anwendungsbeschreibung

Application Notes: standard: 10000 pg/mL. Test principle: This assay employs the competitive inhibition enzyme immunoassay technique. The microtiter plate provided in this kit has been pre-coated with Mouse CTXI protein. Standards or samples are added to the appropriate microtiter plate wells then with a biotin-conjugated antibody specific to Mouse CTXI. Next, Avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. After TMB substrate solution is added. The enzyme-substrate reaction is terminated by the addition of sulphuric acid solution and the color change is measured spectrophotometrically at a wavelength of 450nm 10nm. The concentration of Mouse CTXI in the samples is then determined by comparing the OD of the samples to the standard curve