

Bitte beachten Sie: Dieses Dokument wurde automatisch erstellt und ist kein Ersatz für das Originaldokument des Herstellers.

Product Datasheet

Human Major Histocompatibility Complex Class I G (MHCG) ELISA Kit BYT-ORB776093

Artikelname	Human Major Histocompatibility Complex Class I G (MHCG) ELISA Kit
Artikelnummer	BYT-ORB776093
Hersteller Artikelnummer	orb776093
Alternativnummer	BYT-ORB776093-48, BYT-ORB776093-96
Hersteller	Biorbyt
Kategorie	Kits/Assays
Applikation	ELISA
Spezies Reaktivität	Human
Produktbeschreibung	The test principle applied in this kit is Sandwich enzyme immunoassay. The microtiter plate provided in this kit has been pre-coated with an antibody specific to Major Histocompatibility Complex Class I G(MHCG)....
Konzentration	24 ng/mL
Detektionsbereich	0.38-24 ng/mL
Sensitivitaet	0.17 ng/mL
UniProt	P17693
Proben	serum, plasma, tissue homogenates and other biological fluids

Anwendungsbeschreibung	<p>Application Notes: standard: 24 ng/mL. Test principle: The test principle applied in this kit is Sandwich enzyme immunoassay. The microtiter plate provided in this kit has been pre-coated with an antibody specific to Human MHCG. Standards or samples are added to the appropriate microtiter plate wells then with a biotin-conjugated antibody specific to Human MHCG. Next, Avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. After TMB substrate solution is added, only those wells that contain Human MHCG, biotin-conjugated antibody and enzyme-conjugated Avidin will exhibit a change in color. The enzyme-substrate reaction is terminated by the addition of sulphuric acid solution and the color change is measured spectrophotometrically at a wavelength of 450nm 10nm. The concentration of Human MHCG in the samples is then determined by comparing the OD of the samples to the standard curve</p>
------------------------	---