

Bitte beachten Sie: Dieses Dokument wurde automatisch erstellt und ist kein Ersatz für das Originaldokument des Herstellers.

Product Datasheet

Rabbit Selectin, Endothelium (SELE) ELISA Kit BYT-ORB776346

Artikelname	Rabbit Selectin, Endothelium (SELE) ELISA Kit
Artikelnummer	BYT-ORB776346
Hersteller Artikelnummer	orb776346
Alternativnummer	BYT-ORB776346-48, BYT-ORB776346-96
Hersteller	Biorbyt
Kategorie	Kits/Assays
Applikation	ELISA
Spezies Reaktivität	Rabbit
Produktbeschreibung	The test principle applied in this kit is Sandwich enzyme immunoassay. The microtiter plate provided in this kit has been pre-coated with an antibody specific to Selectin, Endothelium(SELE)....
Konzentration	400 ng/mL
Detektionsbereich	6.25-400 ng/mL
Sensitivitaet	2.24 ng/mL
UniProt	P27113
Proben	Serum, plasma and other biological fluids.

Anwendungsbeschreibung

Application Notes: standard: 400 ng/mL. Test principle: The test principle applied in this kit is Sandwich enzyme immunoassay. The microtiter plate provided in this kit has been pre-coated with an antibody specific to Rabbit SELE. Standards or samples are added to the appropriate microtiter plate wells then with a biotin-conjugated antibody specific to Rabbit SELE. Next, Avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. After TMB substrate solution is added, only those wells that contain Rabbit SELE, biotin-conjugated antibody and enzyme-conjugated Avidin will exhibit a change in color. The enzyme-substrate reaction is terminated by the addition of sulphuric acid solution and the color change is measured spectrophotometrically at a wavelength of 450nm 10nm. The concentration of Rabbit SELE in the samples is then determined by comparing the OD of the samples to the standard curve