

Bitte beachten Sie: Dieses Dokument wurde automatisch erstellt und ist kein Ersatz für das Originaldokument des Herstellers.

## Product Datasheet

### Mouse Plasminogen Activator, Urokinase Receptor (uPAR) ELISA Kit BYT-ORB776377

Artikelname	Mouse Plasminogen Activator, Urokinase Receptor (uPAR) ELISA Kit
Artikelnummer	BYT-ORB776377
Hersteller Artikelnummer	orb776377
Alternativnummer	BYT-ORB776377-48, BYT-ORB776377-96
Hersteller	Biorbyt
Kategorie	Kits/Assays
Applikation	ELISA
Spezies Reaktivität	Mouse
Produktbeschreibung	The test principle applied in this kit is Sandwich enzyme immunoassay. The microtiter plate provided in this kit has been pre-coated with an antibody specific to Plasminogen Activator, Urokinase Receptor(uPAR)....
Konzentration	5000 pg/mL
Detektionsbereich	78.13-5000 pg/mL
Sensitivitaet	34 pg/mL
Proben	Serum, plasma, tissue homogenates and other biological fluids

Anwendungsbeschreibung

Application Notes: standard: 5000 pg/mL. Test principle: The test principle applied in this kit is Sandwich enzyme immunoassay. The microtiter plate provided in this kit has been pre-coated with an antibody specific to Mouse uPAR. Standards or samples are added to the appropriate microtiter plate wells then with a biotin-conjugated antibody specific to Mouse uPAR. Next, Avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. After TMB substrate solution is added, only those wells that contain Mouse uPAR, biotin-conjugated antibody and enzyme-conjugated Avidin will exhibit a change in color. The enzyme-substrate reaction is terminated by the addition of sulphuric acid solution and the color change is measured spectrophotometrically at a wavelength of 450nm 10nm. The concentration of Mouse uPAR in the samples is then determined by comparing the OD of the samples to the standard curve