

Bitte beachten Sie: Dieses Dokument wurde automatisch erstellt und ist kein Ersatz für das Originaldokument des Herstellers.

Product Datasheet

Human Procollagen II N-Terminal Propeptide (PIINP) ELISA Kit BYT-ORB776437

Artikelname	Human Procollagen II N-Terminal Propeptide (PIINP) ELISA Kit
Artikelnummer	BYT-ORB776437
Hersteller Artikelnummer	orb776437
Alternativnummer	BYT-ORB776437-48, BYT-ORB776437-96
Hersteller	Biorbyt
Kategorie	Kits/Assays
Applikation	ELISA
Spezies Reaktivität	Human
Produktbeschreibung	The test principle applied in this kit is Sandwich enzyme immunoassay. The microtiter plate provided in this kit has been pre-coated with an antibody specific to Procollagen II N-Terminal Propeptide(PIINP)....
Konzentration	6000 pg/mL
Detektionsbereich	93.75-6000 pg/mL
Sensitivitaet	38.33 pg/mL
Proben	Serum, plasma, tissue homogenates and other biological fluids

Anwendungsbeschreibung

Application Notes: standard: 6000 pg/mL. Test principle: The test principle applied in this kit is Sandwich enzyme immunoassay. The microtiter plate provided in this kit has been pre-coated with an antibody specific to Human PIINP. Standards or samples are added to the appropriate microtiter plate wells then with a biotin-conjugated antibody specific to Human PIINP. Next, Avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. After TMB substrate solution is added, only those wells that contain Human PIINP, biotin-conjugated antibody and enzyme-conjugated Avidin will exhibit a change in color. The enzyme-substrate reaction is terminated by the addition of sulphuric acid solution and the color change is measured spectrophotometrically at a wavelength of 450nm 10nm. The concentration of Human PIINP in the samples is then determined by comparing the OD of the samples to the standard curve