

Bitte beachten Sie: Dieses Dokument wurde automatisch erstellt und ist kein Ersatz für das Originaldokument des Herstellers.

Product Datasheet

Human NitricOxide Synthase 2, Inducible (NOS2/iNOS) ELISA Kit BYT-ORB776507

Artikelname	Human NitricOxide Synthase 2, Inducible (NOS2/iNOS) ELISA Kit
Artikelnummer	BYT-ORB776507
Hersteller Artikelnummer	orb776507
Alternativnummer	BYT-ORB776507-48, BYT-ORB776507-96
Hersteller	Biorbyt
Kategorie	Kits/Assays
Applikation	ELISA
Spezies Reaktivität	Human
Produktbeschreibung	The test principle applied in this kit is Sandwich enzyme immunoassay. The microtiter plate provided in this kit has been pre-coated with an antibody specific to Nitric Oxide Synthase 2, Inducible(NOS2)....
Konzentration	10 ng/mL
Detektionsbereich	0.16-10 ng/mL
Sensitivitaet	0.054 ng/mL
UniProt	P35228
Proben	serum, plasma, tissue homogenates, cell lysates, cell culture supernates and other biological fluids

Anwendungsbeschreibung

Application Notes: standard: 10 ng/mL. Test principle: The test principle applied in this kit is Sandwich enzyme immunoassay. The microtiter plate provided in this kit has been pre-coated with an antibody specific to Human NOS2/iNOS. Standards or samples are added to the appropriate microtiter plate wells then with a biotin-conjugated antibody specific to Human NOS2/iNOS. Next, Avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. After TMB substrate solution is added, only those wells that contain Human NOS2/iNOS, biotin-conjugated antibody and enzyme-conjugated Avidin will exhibit a change in color. The enzyme-substrate reaction is terminated by the addition of sulphuric acid solution and the color change is measured spectrophotometrically at a wavelength of 450nm 10nm. The concentration of Human NOS2/iNOS in the samples is then determined by comparing the OD of the samples to the standard curve