

Bitte beachten Sie: Dieses Dokument wurde automatisch erstellt und ist kein Ersatz für das Originaldokument des Herstellers.

Product Datasheet

Human Inducible T-Cell Co Stimulator (ICOS) ELISA Kit BYT-ORB776998

Artikelname	Human Inducible T-Cell Co Stimulator (ICOS) ELISA Kit
Artikelnummer	BYT-ORB776998
Hersteller Artikelnummer	orb776998
Alternativnummer	BYT-ORB776998-48, BYT-ORB776998-96
Hersteller	Biorbyt
Kategorie	Kits/Assays
Applikation	ELISA
Spezies Reaktivität	Human
Produktbeschreibung	The test principle applied in this kit is Sandwich enzyme immunoassay. The microtiter plate provided in this kit has been pre-coated with an antibody specific to Inducible T-Cell Co Stimulator(ICOS)....
Konzentration	1000 pg/mL
Detektionsbereich	15.63-1000 pg/mL
Sensitivitaet	5.6 pg/mL
UniProt	Q9Y6W8
Proben	serum, plasma, urine, cell lysates and other biological fluids

Anwendungsbeschreibung

Application Notes: standard: 1000 pg/mL. Test principle: The test principle applied in this kit is Sandwich enzyme immunoassay. The microtiter plate provided in this kit has been pre-coated with an antibody specific to Human ICOS. Standards or samples are added to the appropriate microtiter plate wells then with a biotin-conjugated antibody specific to Human ICOS. Next, Avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. After TMB substrate solution is added, only those wells that contain Human ICOS, biotin-conjugated antibody and enzyme-conjugated Avidin will exhibit a change in color. The enzyme-substrate reaction is terminated by the addition of sulphuric acid solution and the color change is measured spectrophotometrically at a wavelength of 450nm 10nm. The concentration of Human ICOS in the samples is then determined by comparing the OD of the samples to the standard curve