

Bitte beachten Sie: Dieses Dokument wurde automatisch erstellt und ist kein Ersatz für das Originaldokument des Herstellers.

Product Datasheet

Human Interferon Gamma Inducible Protein 30 (IFI30) ELISA Kit BYT-ORB777067

Artikelname	Human Interferon Gamma Inducible Protein 30 (IFI30) ELISA Kit
Artikelnummer	BYT-ORB777067
Hersteller Artikelnummer	orb777067
Alternativnummer	BYT-ORB777067-48, BYT-ORB777067-96
Hersteller	Biorbyt
Kategorie	Kits/Assays
Applikation	ELISA
Spezies Reaktivität	Human
Produktbeschreibung	The test principle applied in this kit is Sandwich enzyme immunoassay. The microtiter plate provided in this kit has been pre-coated with an antibody specific to Interferon Gamma Inducible Protein 30(IFI30)....
Konzentration	10 ng/mL
Detektionsbereich	0.16-10 ng/mL
Sensitivitaet	0.056 ng/mL
UniProt	P13284
Proben	tissue homogenates, cell lysates and other biological fluids

Anwendungsbeschreibung

Application Notes: standard: 10 ng/mL. Test principle: The test principle applied in this kit is Sandwich enzyme immunoassay. The microtiter plate provided in this kit has been pre-coated with an antibody specific to Human IFI30. Standards or samples are added to the appropriate microtiter plate wells then with a biotin-conjugated antibody specific to Human IFI30. Next, Avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. After TMB substrate solution is added, only those wells that contain Human IFI30, biotin-conjugated antibody and enzyme-conjugated Avidin will exhibit a change in color. The enzyme-substrate reaction is terminated by the addition of sulphuric acid solution and the color change is measured spectrophotometrically at a wavelength of 450nm 10nm. The concentration of Human IFI30 in the samples is then determined by comparing the OD of the samples to the standard curve