

Bitte beachten Sie: Dieses Dokument wurde automatisch erstellt und ist kein Ersatz für das Originaldokument des Herstellers.

Product Datasheet

Human Collagen Type IX Alpha 3 (COL9a3) ELISA Kit BYT-ORB777197

Artikelname	Human Collagen Type IX Alpha 3 (COL9a3) ELISA Kit
Artikelnummer	BYT-ORB777197
Hersteller Artikelnummer	orb777197
Alternativnummer	BYT-ORB777197-48, BYT-ORB777197-96
Hersteller	Biorbyt
Kategorie	Kits/Assays
Applikation	ELISA
Spezies Reaktivität	Human
Produktbeschreibung	The test principle applied in this kit is Sandwich enzyme immunoassay. The microtiter plate provided in this kit has been pre-coated with an antibody specific to Collagen Type IX Alpha 3(COL9a3)....
Konzentration	50 ng/mL
Detektionsbereich	0.79-50 ng/mL
Sensitivitaet	0.33 ng/mL
UniProt	Q14050
Proben	Tissue homogenates and other biological fluids.

Anwendungsbeschreibung

Application Notes: standard: 50 ng/mL. Test principle: The test principle applied in this kit is Sandwich enzyme immunoassay. The microtiter plate provided in this kit has been pre-coated with an antibody specific to Human COL9a3. Standards or samples are added to the appropriate microtiter plate wells then with a biotin-conjugated antibody specific to Human COL9a3. Next, Avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. After TMB substrate solution is added, only those wells that contain Human COL9a3, biotin-conjugated antibody and enzyme-conjugated Avidin will exhibit a change in color. The enzyme-substrate reaction is terminated by the addition of sulphuric acid solution and the color change is measured spectrophotometrically at a wavelength of 450nm 10nm. The concentration of Human COL9a3 in the samples is then determined by comparing the OD of the samples to the standard curve