BIOZOL
Diagnostica Vertrieb GmbH, Oehleckerring 11-13
22419 Hamburg, Germany
Telephone: +49 (0)89 3799666-6 | **Fax:** +49 (0)89 3799666-99
E-Mail: info@biozol.de

Bitte beachten Sie: Dieses Dokument wurde automatisch erstellt und ist kein Ersatz für das Originaldokument des Herstellers.

Product Datasheet

Mouse Endosulfine Alpha (ENSa) ELISA Kit
BYT-ORB780061

Artikelname	Mouse Endosulfine Alpha (ENSa) ELISA Kit
Artikelnummer	BYT-ORB780061
Hersteller Artikelnummer	orb780061
Alternativnummer	BYT-ORB780061-48, BYT-ORB780061-96
Hersteller	Biorbyt
Kategorie	Kits/Assays
Applikation	ELISA
Spezies Reaktivität	Mouse
Produktbeschreibung	The test principle applied in this kit is Sandwich enzyme immunoassay. The microtiter plate provided in this kit has been pre-coated with an antibody specific to Endosulfine Alpha(ENSa)....
Konzentration	10 ng/mL
Detektionsbereich	0.16-10 ng/mL
Sensitivitaet	0.057 ng/mL
UniProt	P60840
Proben	tissue homogenates, cell lysates and other biological fluids

Anwendungsbeschreibung	Application Notes: standard: 10 ng/mL. Test principle: The test principle applied in this kit is Sandwich enzyme immunoassay. The microtiter plate provided in this kit has been pre-coated with an antibody specific to Mouse ENSa. Standards or samples are added to the appropriate microtiter plate wells then with a biotin-conjugated antibody specific to Mouse ENSa. Next, Avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. After TMB substrate solution is added, only those wells that contain Mouse ENSa, biotin-conjugated antibody and enzyme-conjugated Avidin will exhibit a change in color. The enzyme-substrate reaction is terminated by the addition of sulphuric acid solution and the color change is measured spectrophotometrically at a wavelength of 450nm 10nm. The concentration of Mouse ENSa in the samples is then determined by comparing the OD of the samples to the standard curve