

Bitte beachten Sie: Dieses Dokument wurde automatisch erstellt und ist kein Ersatz für das Originaldokument des Herstellers.

Product Datasheet

Mouse G Protein Coupled Estrogen Receptor 1 (GPER) ELISA Kit BYT-ORB780776

| | |
|--------------------------|-------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|
| Artikelname | Mouse G Protein Coupled Estrogen Receptor 1 (GPER) ELISA Kit |
| Artikelnummer | BYT-ORB780776 |
| Hersteller Artikelnummer | orb780776 |
| Alternativnummer | BYT-ORB780776-48, BYT-ORB780776-96 |
| Hersteller | Biorbyt |
| Kategorie | Kits/Assays |
| Applikation | ELISA |
| Spezies Reaktivität | Mouse |
| Produktbeschreibung | The test principle applied in this kit is Sandwich enzyme immunoassay. The microtiter plate provided in this kit has been pre-coated with an antibody specific to G Protein Coupled Estrogen Receptor 1(GPER).... |
| Konzentration | 10 ng/mL |
| Detektionsbereich | 0.16-10 ng/mL |
| Sensitivitaet | 0.053 ng/mL |
| UniProt | Q8BMP4 |
| Proben | Serum, plasma, tissue homogenates and other biological fluids. |

Anwendungsbeschreibung

Application Notes: standard: 10 ng/mL. Test principle: The test principle applied in this kit is Sandwich enzyme immunoassay. The microtiter plate provided in this kit has been pre-coated with an antibody specific to Mouse GPER. Standards or samples are added to the appropriate microtiter plate wells then with a biotin-conjugated antibody specific to Mouse GPER. Next, Avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. After TMB substrate solution is added, only those wells that contain Mouse GPER, biotin-conjugated antibody and enzyme-conjugated Avidin will exhibit a change in color. The enzyme-substrate reaction is terminated by the addition of sulphuric acid solution and the color change is measured spectrophotometrically at a wavelength of 450nm 10nm. The concentration of Mouse GPER in the samples is then determined by comparing the OD of the samples to the standard curve