

Bitte beachten Sie: Dieses Dokument wurde automatisch erstellt und ist kein Ersatz für das Originaldokument des Herstellers.

## Product Datasheet

### Mouse Urocortin 1 (UCN1) ELISA Kit BYT-ORB781502

Artikelname	Mouse Urocortin 1 (UCN1) ELISA Kit
Artikelnummer	BYT-ORB781502
Hersteller Artikelnummer	orb781502
Alternativnummer	BYT-ORB781502-48, BYT-ORB781502-96
Hersteller	Biorbyt
Kategorie	Kits/Assays
Applikation	ELISA
Spezies Reaktivität	Mouse
Produktbeschreibung	The test principle applied in this kit is Sandwich enzyme immunoassay. The microtiter plate provided in this kit has been pre-coated with an antibody specific to Urocortin 1(UCN1)....
Konzentration	500 pg/mL
Detektionsbereich	7.82-500 pg/mL
Sensitivitaet	3.3 pg/mL
UniProt	<a href="#">P81615</a>
Proben	serum, plasma, tissue homogenates, cell lysates, cell culture supernates and other biological fluids

Anwendungsbeschreibung

Application Notes: standard: 500 pg/mL. Test principle: The test principle applied in this kit is Sandwich enzyme immunoassay. The microtiter plate provided in this kit has been pre-coated with an antibody specific to Mouse UCN1. Standards or samples are added to the appropriate microtiter plate wells then with a biotin-conjugated antibody specific to Mouse UCN1. Next, Avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. After TMB substrate solution is added, only those wells that contain Mouse UCN1, biotin-conjugated antibody and enzyme-conjugated Avidin will exhibit a change in color. The enzyme-substrate reaction is terminated by the addition of sulphuric acid solution and the color change is measured spectrophotometrically at a wavelength of 450nm 10nm. The concentration of Mouse UCN1 in the samples is then determined by comparing the OD of the samples to the standard curve