

Bitte beachten Sie: Dieses Dokument wurde automatisch erstellt und ist kein Ersatz für das Originaldokument des Herstellers.

## Product Datasheet

### Human Anti-Glutamate Receptor, Ionotropic, N-Methyl-D-Aspartate 2A Antibody (Anti-GRIN2A) ELISA Kit BYT-ORB782077

Artikelname	Human Anti-Glutamate Receptor, Ionotropic, N-Methyl-D-Aspartate 2A Antibody (Anti-GRIN2A) ELISA Kit
Artikelnummer	BYT-ORB782077
Hersteller Artikelnummer	orb782077
Alternativnummer	BYT-ORB782077-48, BYT-ORB782077-96
Hersteller	Biorbyt
Kategorie	Kits/Assays
Applikation	ELISA
Spezies Reaktivität	Human
Produktbeschreibung	This assay employs the competitive inhibition enzyme immunoassay technique. The microtiter plate provided in this kit has been pre-coated with Anti-Glutamate Receptor, Ionotropic, N-Methyl-D-Aspartate 2A Antibody(Anti-GRIN2A) protein. Standards or sa...
Konzentration	200 ng/mL
Detektionsbereich	3.13-200 ng/mL
Sensitivitaet	1.3 ng/mL
Proben	serum, plasma and other biological fluids

Anwendungsbeschreibung

Application Notes: standard: 200 ng/mL. Test principle: This assay employs the competitive inhibition enzyme immunoassay technique. The microtiter plate provided in this kit has been pre-coated with Human Anti-GRIN2A protein. Standards or samples are added to the appropriate microtiter plate wells then with a biotin-conjugated antibody specific to Human Anti-GRIN2A. Next, Avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. After TMB substrate solution is added. The enzyme-substrate reaction is terminated by the addition of sulphuric acid solution and the color change is measured spectrophotometrically at a wavelength of 450nm 10nm. The concentration of Human Anti-GRIN2A in the samples is then determined by comparing the OD of the samples to the standard curve