

Bitte beachten Sie: Dieses Dokument wurde automatisch erstellt und ist kein Ersatz für das Originaldokument des Herstellers.

Product Datasheet

Dog Cross Linked C-Telopeptide of Type III Collagen (CTXIII) ELISA Kit BYT-ORB782390

Artikelname	Dog Cross Linked C-Telopeptide of Type III Collagen (CTXIII) ELISA Kit
Artikelnummer	BYT-ORB782390
Hersteller Artikelnummer	orb782390
Alternativnummer	BYT-ORB782390-48, BYT-ORB782390-96
Hersteller	Biorbyt
Kategorie	Kits/Assays
Applikation	ELISA
Spezies Reaktivität	Canine
Produktbeschreibung	This assay employs the competitive inhibition enzyme immunoassay technique. The microtiter plate provided in this kit has been pre-coated with Cross Linked C-Telopeptide Of Type III Collagen(CTXIII) protein. Standards or samples are added to the appr...
Konzentration	10000 pg/mL
Detektionsbereich	156.25-10000 pg/mL
Sensitivitaet	45.8 pg/mL
Proben	Serum, plasma and other biological fluids

Anwendungsbeschreibung

Application Notes: standard: 10000 pg/mL. Test principle: This assay employs the competitive inhibition enzyme immunoassay technique. The microtiter plate provided in this kit has been pre-coated with Dog CTXIII protein. Standards or samples are added to the appropriate microtiter plate wells then with a biotin-conjugated antibody specific to Dog CTXIII. Next, Avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. After TMB substrate solution is added. The enzyme-substrate reaction is terminated by the addition of sulphuric acid solution and the color change is measured spectrophotometrically at a wavelength of 450nm 10nm. The concentration of Dog CTXIII in the samples is then determined by comparing the OD of the samples to the standard curve