

Bitte beachten Sie: Dieses Dokument wurde automatisch erstellt und ist kein Ersatz für das Originaldokument des Herstellers.

Product Datasheet

Mouse Collagen Type II Alpha 1 (COL2a1) ELISA Kit BYT-ORB782421

| | |
|--------------------------|--|
| Artikelname | Mouse Collagen Type II Alpha 1 (COL2a1) ELISA Kit |
| Artikelnummer | BYT-ORB782421 |
| Hersteller Artikelnummer | orb782421 |
| Alternativnummer | BYT-ORB782421-48, BYT-ORB782421-96 |
| Hersteller | Biorbyt |
| Kategorie | Kits/Assays |
| Applikation | ELISA |
| Spezies Reaktivität | Mouse |
| Produktbeschreibung | The test principle applied in this kit is Sandwich enzyme immunoassay. The microtiter plate provided in this kit has been pre-coated with an antibody specific to Collagen Type II Alpha 1(COL2a1).... |
| Konzentration | 10 ng/mL |
| Detektionsbereich | 0.16-10 ng/mL |
| Sensitivitaet | 0.059 ng/mL |
| UniProt | P28481 |
| Proben | Serum, plasma and other biological fluids. |

Anwendungsbeschreibung

Application Notes: standard: 10 ng/mL. Test principle: The test principle applied in this kit is Sandwich enzyme immunoassay. The microtiter plate provided in this kit has been pre-coated with an antibody specific to Mouse COL2a1. Standards or samples are added to the appropriate microtiter plate wells then with a biotin-conjugated antibody specific to Mouse COL2a1. Next, Avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. After TMB substrate solution is added, only those wells that contain Mouse COL2a1, biotin-conjugated antibody and enzyme-conjugated Avidin will exhibit a change in color. The enzyme-substrate reaction is terminated by the addition of sulphuric acid solution and the color change is measured spectrophotometrically at a wavelength of 450nm 10nm. The concentration of Mouse COL2a1 in the samples is then determined by comparing the OD of the samples to the standard curve