

Bitte beachten Sie: Dieses Dokument wurde automatisch erstellt und ist kein Ersatz für das Originaldokument des Herstellers.

Product Datasheet

Human IsocitRate Dehydrogenase (ICD) ELISA Kit BYT-ORB782469

| | |
|--------------------------|-----------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|
| Artikelname | Human IsocitRate Dehydrogenase (ICD) ELISA Kit |
| Artikelnummer | BYT-ORB782469 |
| Hersteller Artikelnummer | orb782469 |
| Alternativnummer | BYT-ORB782469-48, BYT-ORB782469-96 |
| Hersteller | Biorbyt |
| Kategorie | Kits/Assays |
| Applikation | ELISA |
| Spezies Reaktivität | Human |
| Produktbeschreibung | The test principle applied in this kit is Sandwich enzyme immunoassay. The microtiter plate provided in this kit has been pre-coated with an antibody specific to ICD(IsocitRate Dehydrogenase).... |
| Konzentration | 20 ng/mL |
| Detektionsbereich | 0.32-20 ng/mL |
| Sensitivitaet | 0.121 ng/mL |
| UniProt | O75874 |
| Proben | Tissue homogenates, cell lysates and other biological fluids |

Anwendungsbeschreibung

Application Notes: standard: 20 ng/mL. Test principle: The test principle applied in this kit is Sandwich enzyme immunoassay. The microtiter plate provided in this kit has been pre-coated with an antibody specific to Human ICD. Standards or samples are added to the appropriate microtiter plate wells then with a biotin-conjugated antibody specific to Human ICD. Next, Avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. After TMB substrate solution is added, only those wells that contain Human ICD, biotin-conjugated antibody and enzyme-conjugated Avidin will exhibit a change in color. The enzyme-substrate reaction is terminated by the addition of sulphuric acid solution and the color change is measured spectrophotometrically at a wavelength of 450nm 10nm. The concentration of Human ICD in the samples is then determined by comparing the OD of the samples to the standard curve