

Bitte beachten Sie: Dieses Dokument wurde automatisch erstellt und ist kein Ersatz für das Originaldokument des Herstellers.

Product Datasheet

Human Circulating Histone H3 ELISA Kit BYT-ORB782537

Artikelname	Human Circulating Histone H3 ELISA Kit
Artikelnummer	BYT-ORB782537
Hersteller Artikelnummer	orb782537
Alternativnummer	BYT-ORB782537-48, BYT-ORB782537-96
Hersteller	Biorbyt
Kategorie	Kits/Assays
Applikation	ELISA
Spezies Reaktivität	Human
Produktbeschreibung	The test principle applied in this kit is Sandwich enzyme immunoassay. The microtiter plate provided in this kit has been pre-coated with an antibody specific to Circulating Histone H3....
Konzentration	100 ng/mL
Detektionsbereich	1.57-100 ng/mL
Sensitivitaet	0.62 ng/mL
Proben	serum, plasma and other biological fluids

Anwendungsbeschreibung

Application Notes: standard: 100 ng/mL. Test principle: The test principle applied in this kit is Sandwich enzyme immunoassay. The microtiter plate provided in this kit has been pre-coated with an antibody specific to Human Circulating Histone H3. Standards or samples are added to the appropriate microtiter plate wells then with a biotin-conjugated antibody specific to Human Circulating Histone H3. Next, Avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. After TMB substrate solution is added, only those wells that contain Human Circulating Histone H3, biotin-conjugated antibody and enzyme-conjugated Avidin will exhibit a change in color. The enzyme-substrate reaction is terminated by the addition of sulphuric acid solution and the color change is measured spectrophotometrically at a wavelength of 450nm 10nm. The concentration of Human Circulating Histone H3 in the samples is then determined by comparing the OD of the samples to the standard curve