

Bitte beachten Sie: Dieses Dokument wurde automatisch erstellt und ist kein Ersatz für das Originaldokument des Herstellers.

Product Datasheet

Mouse Alpha-ketoglutaRate-dependent Dioxygenase FTO (FTO) ELISA Kit BYT-ORB782632

Artikelname	Mouse Alpha-ketoglutaRate-dependent Dioxygenase FTO (FTO) ELISA Kit
Artikelnummer	BYT-ORB782632
Hersteller Artikelnummer	orb782632
Alternativnummer	BYT-ORB782632-48, BYT-ORB782632-96
Hersteller	Biorbyt
Kategorie	Kits/Assays
Applikation	ELISA
Spezies Reaktivität	Mouse
Produktbeschreibung	The test principle applied in this kit is Sandwich enzyme immunoassay. The microtiter plate provided in this kit has been pre-coated with an antibody specific to FTO(Alpha-ketoglutaRate-dependent dioxygenase FTO)....
Konzentration	10 ng/mL
Detektionsbereich	0.16-10 ng/mL
Sensitivitaet	0.06 ng/mL
UniProt	Q8BGW1
Proben	serum, plasma and other biological fluids

Anwendungsbeschreibung

Application Notes: standard: 10 ng/mL. Test principle: The test principle applied in this kit is Sandwich enzyme immunoassay. The microtiter plate provided in this kit has been pre-coated with an antibody specific to Mouse FTO. Standards or samples are added to the appropriate microtiter plate wells then with a biotin-conjugated antibody specific to Mouse FTO. Next, Avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. After TMB substrate solution is added, only those wells that contain Mouse FTO, biotin-conjugated antibody and enzyme-conjugated Avidin will exhibit a change in color. The enzyme-substrate reaction is terminated by the addition of sulphuric acid solution and the color change is measured spectrophotometrically at a wavelength of 450nm 10nm. The concentration of Mouse FTO in the samples is then determined by comparing the OD of the samples to the standard curve