

Bitte beachten Sie: Dieses Dokument wurde automatisch erstellt und ist kein Ersatz für das Originaldokument des Herstellers.

Product Datasheet

Rat Cluster of Differentiation 34 (CD34) ELISA Kit BYT-ORB782707

Artikelname	Rat Cluster of Differentiation 34 (CD34) ELISA Kit
Artikelnummer	BYT-ORB782707
Hersteller Artikelnummer	orb782707
Alternativnummer	BYT-ORB782707-48, BYT-ORB782707-96
Hersteller	Biorbyt
Kategorie	Kits/Assays
Applikation	ELISA
Spezies Reaktivität	Rat
Produktbeschreibung	The test principle applied in this kit is Sandwich enzyme immunoassay. The microtiter plate provided in this kit has been pre-coated with an antibody specific to Cluster Of Differentiation 34(CD34)....
Konzentration	50 ng/mL
Detektionsbereich	0.79-50 ng/mL
Sensitivitaet	0.37 ng/mL
Proben	tissue homogenates, cell lysates, cell culture supernates and other biological fluids

Anwendungsbeschreibung

Application Notes: standard: 50 ng/mL. Test principle: The test principle applied in this kit is Sandwich enzyme immunoassay. The microtiter plate provided in this kit has been pre-coated with an antibody specific to Rat CD34. Standards or samples are added to the appropriate microtiter plate wells then with a biotin-conjugated antibody specific to Rat CD34. Next, Avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. After TMB substrate solution is added, only those wells that contain Rat CD34, biotin-conjugated antibody and enzyme-conjugated Avidin will exhibit a change in color. The enzyme-substrate reaction is terminated by the addition of sulphuric acid solution and the color change is measured spectrophotometrically at a wavelength of 450nm 10nm. The concentration of Rat CD34 in the samples is then determined by comparing the OD of the samples to the standard curve