

Bitte beachten Sie: Dieses Dokument wurde automatisch erstellt und ist kein Ersatz für das Originaldokument des Herstellers.

Product Datasheet

Human SARS-CoV-2 Nucleocapsid ELISA Kit BYT-ORB867010

Artikelname	Human SARS-CoV-2 Nucleocapsid ELISA Kit
Artikelnummer	BYT-ORB867010
Hersteller Artikelnummer	orb867010
Alternativnummer	BYT-ORB867010-48,BYT-ORB867010-96
Hersteller	Biorbyt
Kategorie	Kits/Assays
Applikation	ELISA
Spezies Reaktivität	Human
Produktbeschreibung	Sandwich Human SARS-CoV-2 Nucleocapsid ELISA Kit...
Konzentration	500 pg/mL
Detektionsbereich	7.82-500 pg/mL
Sensitivitaet	3.2 pg/mL
Proben	serum, plasma, saliva, tissue homogenates, cell lysates, cell culture supernates and other biological fluids

Anwendungsbeschreibung	Application Notes: standard: 500 pg/mL. Test principle: The test principle applied in this kit is Sandwich enzyme immunoassay. The microtiter plate provided in this kit has been pre-coated with an antibody specific to Human SARS-CoV-2 Nucleocapsid. Standards or samples are added to the appropriate microtiter plate wells then with a biotin-conjugated antibody specific to Human SARS-CoV-2 Nucleocapsid. Next, Avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. After TMB substrate solution is added, only those wells that contain Human SARS-CoV-2 Nucleocapsid, biotin-conjugated antibody and enzyme-conjugated Avidin will exhibit a change in color. The enzyme-substrate reaction is terminated by the addition of sulphuric acid solution and the color change is measured spectrophotometrically at a wavelength of 450nm 10nm. The concentration of Human SARS-CoV-2 Nucleocapsid in the samples is then determined by comparing the OD of the samples to the standard curve
------------------------	--