BIOZOL

Diagnostica Vertrieb GmbH, Oehleckerring 11-13

22419 Hamburg, Germany

Telephone: +49 (0)89 3799666-6 | **Fax:** +49 (0)89 3799666-99

E-Mail: info@biozol.de

Bitte beachten Sie: Dieses Dokument wurde automatisch erstellt und ist kein Ersatz für das Originaldokument des Herstellers.

Product Datasheet

Goat IgG anti-Bovine IgG (F(ab)2)-FITC, MinX none

DNA-SEC-182480

Artikelname	Goat IgG anti-Bovine IgG (F(ab)2)-FITC, MinX none
Artikelnummer	DNA-SEC-182480
Hersteller Artikelnummer	SEC-182480
Alternativnummer	DNA-SEC-182480
Hersteller	dianova
Wirt	Goat
Kategorie	Antikörper
Applikation	FLISA,FACS,IF
Spezies Reaktivität	Bovine
Immunogen	Bovine IgG F(ab)2 fragment
Konjugation	FITC
Format	IgG
Spezifität	IgG (F(ab')2)
Minimale Kreuzreaktivität (MinX)	no cross-adsorbtion

Produktbeschreibung	Secreted as part of the adaptive immune response by plasma B cells, immunoglobulin G constitutes 75% of serum immunoglobulins. Immunoglobulin G binds to viruses, bacteria, as well as fungi and facilitates their destruction or neutralization via agglu...
Klonalität	Polyclonal
Isotyp	Ig
Puffer	0.02 M Potassium Phosphate, 0.15 M Sodium Chloride, pH 7.2
Reinheit	This product is an IgG fraction antibody purified from monospecific antiserum by a multi-step process which includes delipidation, salt fractionation and ion exchange chromatography followed by extensive dialysis against the buffer stated above. Assay by immunoelectrophoresis resulted in a single precipitin arc against anti-fluorescein, anti-Goat Serum, Bovine IgG, Bovine IgG F(ab')2 and Bovine Serum. No reaction was observed against Bovine IgG F(c).
Formulierung	Lyophilized
Formel	20 mM K3PO4,150 mM NaCl,pH 7,2,lyophilisate,Azide/BSA free
Target-Kategorie	Bovine
Antibody Type	Secondary Antibody
Application Verdünnung	FLISA Dilution: 1:10,000 - 1:1:50,000, Flow Cytometry Dilution: 1:500 - 1:2,500, IF Microscopy Dilution: 1:1,000 - 1:1:5,000
Anwendungsbeschreibung	Suitable for immunomicroscopy and flow cytometry or FACS analysis as well as other antibody based fluorescent assays requiring extremely low background levels, absence of F(c) mediated binding, lot-to-lot consistency, high titer and specificity. The maximum amount of reagent required to stain 1 x 10E6 cells in flow cytometry is approximately 1.0 µg of antibody. Lesser amounts of reagent may be sufficient for staining. Optimal titers for other applications should be determined by the researcher. As a general guideline dilutions of 1:100 to 1:250 should be suitable for most applications.