

Please note: This document was created automatically and is not a substitute for the manufacturer's original document.

## Product Datasheet

### **Chloride Anion IC Standard from Sodium Chloride at 100 µg/mL in H2O 100 mL HPS-IC-CL-100**

Article Name	Chloride Anion IC Standard from Sodium Chloride at 100 µg/mL in H2O 100 mL
Biozol Catalog Number	HPS-IC-CL-100
Supplier Catalog Number	IC-CL-100
Alternative Catalog Number	HPS-IC-CL-100-100
Manufacturer	High Purity Standards
Category	Biochemikalien
Product Description	Chloride 100 µg/mL in H2O, 12 months expiry date. Traceable to NIST 3182. ISO 9001:2015 certified, ISO/IEC 17025:2017 and ISO 17034:2016 accredited....