

Diagnostica Vertrieb GmbH, Oehleckerring 11-13

22419 Hamburg, Germany

**Telephone:** +49 (0)89 3799666-6 | **Fax:** +49 (0)89 3799666-99

E-Mail: info@biozol.de

Please note: This document was created automatically and is not a substitute for the manufacturer's original document.

## **Product Datasheet**

## Phosphorus Anion IC Standard from Potassium Dihydrogen Phosphate at 100 $\mu g/mL$ in H2O HPS-IC-KP-250

Article Name	Phosphorus Anion IC Standard from Potassium Dihydrogen Phosphate at 100 µg/mL in H2O
Biozol Catalog Number	HPS-IC-KP-250
Supplier Catalog Number	IC-KP-250
Alternative Catalog Number	HPS-IC-KP-250-250
Manufacturer	High Purity Standards
Category	Biochemikalien
Product Description	Phosphorus from KH2PO4100 ug/mL in H2O, 18 months expiry date. Traceable to NIST 3186. ISO 9001:2015 certified, ISO/IEC 17025:2017 and ISO 17034:20166 accredited. Can be used as stock solution for EPA methods 365.1, 365.4, SM-4500-P, -F or -G and 300