



Diagnostica Vertrieb GmbH, Oehleckerring 11-13

22419 Hamburg, Germany

Telephone: +49 (0)89 3799666-6 | Fax: +49 (0)89 3799666-99

E-Mail: info@biozol.de

Please note: This document was created automatically and is not a substitute for the manufacturer's original document.

Product Datasheet

Iron(III) (Fe+3) Speciation Standard at 1,000 μ g/mL in 2% HNO3 HPS-100026-7-500

Article Name	Iron(III) (Fe+3) Speciation Standard at 1,000 μg/mL in 2% HNO3
Biozol Catalog Number	HPS-100026-7-500
Supplier Catalog Number	100026-7-500
Alternative Catalog Number	HPS-100026-7-500-500
Manufacturer	High Purity Standards
Category	Biochemikalien
Product Description	Fe+3 1000 μ g/mL in 2% HNO3 12 months expiry date. Traceable to NIST SRM 3126 for total iron concentration and additionally tested by IC against a second source standard for the speciated form. Certified value based on gravimetric preparation. ISO 900