

Please note: This document was created automatically and is not a substitute for the manufacturer's original document.

## Product Datasheet

### **Mercury (Hg) ICP Standard at 1,000 µg/mL in 4% HCl + 3% HNO<sub>3</sub> 500 mL HPS-100077-12-500**

Article Name	Mercury (Hg) ICP Standard at 1,000 µg/mL in 4% HCl + 3% HNO <sub>3</sub> 500 mL
Biozol Catalog Number	HPS-100077-12-500
Supplier Catalog Number	100077-12-500
Alternative Catalog Number	HPS-100077-12-500-500
Manufacturer	High Purity Standards
Category	Biochemikalien
Product Description	Hg 1000 µg/mL in 4% HCl + 3% HNO <sub>3</sub> , 18 months expiry date. Traceable to NIST SRM 3177. ISO 9001:2015 certified, ISO/IEC 17025:2017 and ISO 17034:2016 accredited....